APPLICATION BRIEF

EA-IRMS: No Intra-sample Memory for δ^{34} S Measurements of Inorganic Sulfur Materials Using the EA IsoLink IRMS System

Christopher Brodie, Oliver Kracht, Alexander Hartwig, Andreas Hilkert, Charles Douthitt Thermo Fisher Scientific, Bremen, Germany

Key Words

 δ^{34} S, Carry-over, EA-IRMS, Intra-sample Memory

Goal

To demonstrate that the EA IsoLink IRMS System is free from intra-sample memory for δ^{34} S measurements by using inorganic international reference materials.



Figure 1. Thermo Scientific EA IsoLink IRMS System.



Introduction

Accurate and precise δ^{34} S measurements by EA-IRMS require that the system can process samples without intra-sample effects, generally referred to as "memory effect" or "carry-over". Any intra-sample memory effect or carry-over will produce inaccurate δ^{34} S values and significantly reduce data reproducibility as well as reducing system productivity and consuming valuable laboratory resources.

This application brief demonstrates that the Thermo Scientific[™] EA IsoLink[™] IRMS System (Figure 1) does not show an intra-sample memory effect on δ^{34} S measurements and that accurate and precise δ^{34} S values are produced.



thermo scientific

Intra-sample performance for $\delta^{34}\textbf{S}$ analysis

Figure 2 illustrates a series of 30 sequential δ^{34} S measurements of international standard reference materials IAEA S1, IAEA S2, IAEA S3, IAEA S05 and IAEA S06. Samples were weighed in amounts of

200–280 µg, which equates to 25–30 µg of sulfur. A 3-point calibration was produced, with an $r^2 = 0.999$. The standard deviation of the mean is 0.19‰ across a 56‰ range (Figure 3). The δ^{34} S precision was $\leq 0.18\%$ for n = 6 showing excellent accuracy (Table 1).









Summary

The analysis of inorganic sulfur materials shows no intra-sample memory or carry-over effect on $\delta^{34}S$ measurements on the EA IsoLink IRMS System, therefore:

- There is no need to discard the first or second sample points from replicate measurements prior to analyzing the data.
- There are no trends in the replicate analysis of individual samples.
- There is no requirement to undertake memory correction.

The data presented in Table 1 are application data and are not warranted because they exceed product specifications. The warranted product specification for δ^{34} S is ±0.2‰ (1 sd) for 50 µg of sulfur measured on Sulfanilamide.

Sample	Weight (mg)	n	µg S	Measured δ³⁴S _{vcb⊺} (‰) [Mean ± 1♂]	Actual δ³⁴S _{vcστ} (‰) [Mean ± 1σ]
IAEA S1 (AgS)	0.202-0.285	6	30.4	0.03 ± 0.08	-0.30 ± 0.20
IAEA S2 (AgS)	0.192-0.219	6	25.8	22.67 ± 0.10	22.70 ± 0.20
IAEA S3 (AgS)	0.191–0.234	6	26.0	-32.30 ± 0.15	-32.30 ± 0.20
IAEA SO5 (BaSO ₄)	0.196-0.268	6	30.4	0.75 ± 0.10	0.50 ± 0.20
IAEA SO6 (BaSO ₄)	0.188-0.272	6	29.2	-34.43 ± 0.18	-34.1 ± 0.20

Table 1. Summary of measured δ^{34} S values and actual δ^{34} S values.

Find out more at **thermofisher.com/EAIsoLink**

For Research Use Only. Not for use in diagnostic procedures. © 2016 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified. AB30357-EN 0716S

Thermo Fisher